

Anti-IgG [Polyclonal]

Catalog No.	Description		
AR050-5R	6 ml of Ready-to-Use Antibody for use with BioGenex Super Sensitive TM Detection Systems OR equivalent detection system		
AR050-10R	10 ml of Ready-to-Use Antibody in a barcode labeled vial for use with BioGenex Super Sensitive TM Detection Systems and i6000 TM Automated Staining Systems		
PU050-UP	1 ml of Concentrated Antibody for use with BioGenex Super Sensitive TM Detection Systems OR equivalent detection system		
PU050-5UP	0.5 ml of Concentrated Antibody for use with BioGenex Super Sensitive TM Detection Systems OR equivalent detection system		
AW050-YCD	Ready-to-Use Antibody in Barcode labeled vial for use on the Xmatrx [®] Elite Staining System, 160 tests		
AW050-50D Ready-to-Use Antibody in Barcode la vial for use on the Xmatrx® Elite Stair System, 50 tests			
AW050-4M	Ready-to-Use Antibody in Barcode labeled vial for use on the NanoVIP® Staining System, 50 tests		

Clone	Species	Ig Class
Polyclonal	Rabbit	N/A

Intended Use

For In Vitro Diagnostic Use. This antibody is designed for the specific localization of IgG in formalin-fixed, paraffinembedded (FFPE) tissue sections. Evaluation must be performed by a qualified pathologist.

Summary and Explanation

The human B-lymphocyte is characterized by the presence of readily detectable surface immunoglobulins. Upto 10 percent of peripheral blood lymphocytes and 68-70 percent of the lymphocytes in lymph nodes are of the B-cell type. When stimulated by antigen and aided by T-helper lymphocytes, Bcells differentiate into plasma cells which are the primary antibody-producing cells. The molecular weight of IgG is 150,000 consisting of two gamma heavy chains and two kappa or lambda light chains. It has been used in the classification of leukemias, plasmacytomas and certain non-Hodgkin's lymphomas.

Storage and Handling

Store at 2-8°C. Fresh dilutions, if required, should be prepared prior to use and are stable and steady for up to one day at room temperature (20-26°C). Diluted antibody preparations can be refrigerated or frozen for extended shelf life.

Principles of the Procedure

Antigen detection by immunohistochemistry (IHC) is a two-step process wherein the primary antibody binds to the antigen of interest and that binding is detected by a chromogen. The primary antibody may be used in IHC using manual techniques or BioGenex Automated Staining System. Positive and negative controls should always be run simultaneously with all patient specimens.

Reagents Provided

Rabbit Polyclonal Antibody IgG is affinity purified and diluted in PBS, pH 7.2, containing 1% BSA and 0.09% sodium azide.

Dilution of Primary Antibody

BioGenex Ready-to-Use antibodies have been optimized for use with the recommended BioGenex Detection System and should not require further dilution.

BioGenex concentrated antibodies must be diluted in accordance with the recommended protocol when used with the recommended BioGenex Detection System.

Recommended Protocol

Refer to the following table for conditions specifically recommended for this antibody. Refer to the BioGenex website for guidance on specific staining protocols or other requirements.

Parameter	BioGenex Recommendations		
Control Tissue	TONSIL tissue as available with Biogenex FB-050P* & FG-050P*		
Recommended Dilution for Concentrated Antibody	1:50-100 in HK156		
Recommended Pretreatment (Manual/i6000)**	EZ-AR2 (HK522-XAK)		
Recommended Pretreatment (Xmatrx & NanoVIP)	EZ-AR2 Elegance (HX032-YCD & HX046- 08XN)		
Antibody Incubation (Manual/i6000)	30-60 Min at RT		
Antibody Incubation (Xmatrx & NanoVIP)	30-60 Min at RT		
Detection System for Manual, Xmatrx, NanoVIP & i6000 systems***	Use BioGenex Two-Step OR One-Step Super Sensitive TM Polymer-HRP IHC Detection System/DAB; see p. 2 for more information		

Category	Antibodies	Revision No.	L
Document No.	932-050P-EN	Release Date	11-May-2022



*FB: positive control micro chamber slides, FG: positive control microscopic slides. Xmatrx & NanoVIP require micro chamber slides.

Pretreatment times will vary based on individual microwave power. *For automation systems (Xmatrx-Elite, NanoVIP & i6000 Diagnostics), refer to the factory protocols provided with the instrument

Detection System	Two-Step HRP Kit	One-Step HRP Kit	Link and Label Kit
M 1	QD440-XAKEN (1000 Test) QD430-XAKEN (1000 Test)	QD630-XAKEN (1000 Test)	QP300- XAKE (1000 Test)
Manual	QD420-YIKEN (500 Test) QD400-60KEN (60 Test)	QD620-XAKEN (500 Test)	QP900- 9LE (500 Test)
Xmatrx - Automation	QD550-YCDEN (200 Test)	QD610-YADEN (200 Test)	N/A
NanoVIP- Automation	QD551-YCDEN (100 Test)	QD611-YADEN (100 Test)	N/A
i6000 - Automation	QD410-YAXEN (200 Test)	QD610-YAXEN (200 Test)	N/A

Precautions

This product contains sodium azide at concentrations of less than 0.1%. Sodium azide is not classified as a hazardous chemical at the product concentrations, but proper handling protocols should be observed. For more information, a Safety Data Sheet (SDS) for sodium azide is available upon request. Dispose of unused reagents according to Local, State and Federal Regulations. Wear suitable Personal Protective Equipment, do not pipette reagents by mouth, and avoid contact of reagents and specimens with skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with copious amounts of water.

Quality Control

Refer to BioGenex detection system documents for guidance on general quality control procedures.

Troubleshooting

Refer to the troubleshooting section in the documentation for BioGenex Detection Systems (or equivalent detection systems) for remedial actions on detection system related issues, or contact BioGenex Technical Support Department at 1-800-421-4149 or support@biogenex.com or your local distributor to report unusual staining.

Expected Results

This antibody stains membrane and cytoplasm in positive cells in formalin-fixed, paraffin embedded tissue sections. An

example image of a tissue section stained with this antibody can be found on the product page on the BioGenex website. Interpretation of the staining result is solely the responsibility of the user. Experimental results should be confirmed by a medically-established diagnostic product or procedure.

Limitations of the Procedure

Improper tissue handling and processing prior to immunostaining can lead to inconsistent results. Variations in embedding and fixation or the nature of the tissue may lead to variations in results. Endogenous peroxidase activity or pseudo peroxidase activity in erythrocytes and tissue biotin may result in non-specific staining based on the detection system employed. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive with horseradish peroxidase systems. Improper counterstaining and mounting may compromise the interpretation of results.

Bibliography

- 1. Grootscholten C et al Arthritis Rheum 48(5):1355-62,
- Taylor, CR. Arch Pathol Lab Med 102:113-121, 1978.
- Ueda, S, et al. Am J Pathol 132:304-318, 1988.

2°C 8°C	Temperature Limitation	IVD	In Vitro Diagnostic Medical Device
\bowtie	Use By Date	LOT	Batch Code
NON	Non-Sterile	i	Consult Instructions for Use
EC REP	Representative in the European Community		Manufacturer

© 2020, BioGenex Laboratories. All rights reserved.

Category	Antibodies	Revision No.	L
Document No.	932-050P-EN	Release Date	11-May-2022