

# Anti-Neurofilament [NE-14]

Catalog No.	Description		
AM073-5M	6 ml of Ready-to-Use Antibody for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems OR equivalent detection system		
AM073-10M	10 ml of Ready-to-Use Antibody in a barcode labeled vial for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems and i6000 <sup>TM</sup> Automated Staining Systems		
MU073-UC	1 ml of Concentrated Antibody for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems OR equivalent detection system		
MU073-5UC	0.5 ml of Concentrated Antibody for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems OR equivalent detection system		
AX073-YCD	Ready-to-Use Antibody in Barcode labeled vial for use on the Xmatrx <sup>®</sup> Elite Staining System, 160 tests		
AX073-50D	Ready-to-Use Antibody in Barcode labeled vial for use on the Xmatrx® Elite Staining System, 50 tests		
AX073-4M	Ready-to-Use Antibody in Barcode labeled		

Clone	Species	Ig Class
NE-14	Mouse	IgG1

#### **Intended Use**

For In Vitro Diagnostic Use. This antibody is designed for the specific localization of neurofilaments in formalin-fixed, paraffinembedded (FFPE) tissue sections. Evaluation must be performed by a qualified pathologist.

### **Summary and Explanation**

Neurofilaments (10 nm diameter) and microtubules (25 nm diameter) comprise the main structural elements of neuronal axons, dendrites, and perikerva. Neurofilaments are composed of three major polypeptides referred to as the neurofilament triplet with approximate molecular weights of 200 kD, 160 kD and 68 kD. This antibody can be used for positive identification of neurons in the central and peripheral nervous systems. In general, co-expression of keratin and neurofilament should be interpreted as indicating neuroendocrine differentiation of a given tissue or neoplasm.

# **Storage and Handling**

Store at 2-8°C. Fresh dilutions, if required, should be prepared prior to use and are stable and steady for up to one day at room temperature (20-26°C). Diluted antibody preparations can be refrigerated or frozen for extended shelf life.

# **Principles of the Procedure**

Antigen detection by immunohistochemistry (IHC) is a two-step process wherein the primary antibody binds to the antigen of interest and that binding is detected by a chromogen. The primary antibody may be used in IHC using manual techniques or BioGenex Automated Staining System. Positive and negative controls should always be run simultaneously with all patient specimens.

# Reagents Provided

Mouse Monoclonal Antibody to neurofilaments is affinity purified and diluted in PBS, pH 7.2, containing 1% BSA and 0.09% sodium azide.

# **Dilution of Primary Antibody**

BioGenex Ready-to-Use antibodies have been optimized for use with the recommended BioGenex Detection System and should not require further dilution.

BioGenex concentrated antibodies must be diluted in accordance with the recommended protocol when used with the recommended BioGenex Detection System.

#### **Recommended Protocol**

Refer to the following table for conditions specifically recommended for this antibody. Refer to the BioGenex website for guidance on specific staining protocols or other requirements.

Parameter	BioGenex Recommendations	
Control Tissue	Nerve as available with Biogenex FB-073M* & FG-073M*	
Recommended Dilution for Concentrated Antibody	1:50-100 in HK156	
Recommended Pretreatment (Manual/i6000)**	EZ-AR2 (HK522-XAK)	
Recommended Pretreatment (Xmatrx & NanoVIP)	EZ-AR2 Elegance (HX032-YCD & HX046-08XN)	
Antibody Incubation (Manual/i6000)	30-60 min at RT	
Antibody Incubation (Xmatrx & NanoVIP)	30-60 min at 25°C	
Detection System for Manual, Xmatrx, NanoVIP & i6000systems***	Use BioGenex Two-Step <b>OR</b> One-Step Super Sensitive <sup>TM</sup> Polymer-HRP IHC Detection System/DAB; see p. 2 for more information	

<sup>\*</sup>FB: positive control micro chamber slides, FG: positive control microscopic slides. Xmatrx & NanoVIP require micro chamber

\*\*Pretreatment times will vary based on individual microwave power. \*\*\*For automation systems (Xmatrx-Elite, NanoVIP & i6000 Diagnostics), refer to the factory protocols provided with the instrument.

Category	Antibodies	Revision No.	J
Document No.	932-073M-EN	Release Date	11-May-2022

Two-Step HRP Kit	One-Step HRP Kit	Link and Label Kit
QD440-XAKEN (1000 Test) QD430-XAKEN (1000 Test)	QD630-XAKEN (1000 Test)	QP300- XAKE (1000 Test)
QD420-YIKEN (500 Test) QD400-60KEN (60 Test)	QD620-XAKEN (500 Test)	QP900-9LE (500 Test)
QD490-YCDEN (200 Test)	QD610-YADEN (200 Test)	N/A
QD551-YCDEN (100 Test)	QD611-YADEN (100 Test)	N/A
QD410-YAXEN (200 Test)	QD610-YAXEN (200 Test)	N/A
	HRP Kit  QD440-XAKEN (1000 Test)  QD430-XAKEN (1000 Test)  QD420-YIKEN (500 Test)  QD400-60KEN (60 Test)  QD490-YCDEN (200 Test)  QD551-YCDEN (100 Test)  QD410-YAXEN	HRP Kit   HRP Kit

#### **Precautions**

This product contains sodium azide at concentrations of less than 0.1%. Sodium azide is not classified as a hazardous chemical at the product concentrations, but proper handling protocols should be observed. For more information, a Safety Data Sheet (SDS) for sodium azide is available upon request. Dispose of unused reagents according to Local, State and Federal Regulations. Wear suitable Personal Protective Equipment, do not pipette reagents by mouth, and avoid contact of reagents and specimens with skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with copious amounts of water.

# **Quality Control**

Refer to BioGenex detection system documents for guidance on general quality control procedures.

# **Troubleshooting**

Refer to the troubleshooting section in the documentation for BioGenex Detection Systems (or equivalent detection systems) for remedial actions on detection system related issues, or contact BioGenex Technical Support Department at 1-800-421-4149 or support@biogenex.com or your local distributor to report unusual staining.

#### **Expected Results**

This antibody stains cytoplasm in positive cells in formalinfixed, paraffin embedded tissue sections. An example image of a tissue section stained with this antibody can be found on the product page on the BioGenex website. Interpretation of the staining result is solely the responsibility of the user. Experimental results should be confirmed by a medicallyestablished diagnostic product or procedure.

# Limitations of the Procedure

Improper tissue handling and processing prior to immunostaining can lead to inconsistent results. Variations in embedding and fixation or the nature of the tissue may lead to variations in results. Endogenous peroxidase activity or pseudo peroxidase activity in erythrocytes and tissue biotin may result in non-specific staining based on the detection system employed. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive with horseradish peroxidase systems. Improper counterstaining and mounting may compromise the interpretation of results.

# **Bibliography**

- Lehto VP, et al. Am J Pathol 110:113-118, 1983.
- Lehto VP, et al. Arch Pathol Lab Med 107:492-494, 1983.
- Kluck P, et al. The Lancet, March 24th, 652-654, 1984.
- Muijen GNP van, et al. Am J Pathol 116:363-369, 1984.
- 5. Muijen GNP van, et al. Hum Pathol 16:590-5.
- 6. Center for Disease Control. Decontamination of Laboratory Sink Drains to Remove Azide Salts. Center for Disease Control Manual Guide--Safety Management, No. CDC-22, Atlanta, Georgia. April 30, 1976.
- Kiernan JA. Histological and Histochemical Methods: Theory and Practice. New York: Pergamon Press 1981.
- 8. Nadji M, Morales AR. Immunoperoxidase, part 1: the techniques and its pitfall. Lab Med 1983: 14:767-770.
- Omata M, Liew CT, Ashcavai M, Peters Rl. Nonimmunologic binding of horseradish peroxidase to hepatitis B surface antigen. A possible source of error in immunohistochemistry. Am J Clin Pathol. May, 1980; 73(5):626-632.
- 10. U.S. Congress. Clinical Laboratory Improvement Amendments of 1988: Final Rule, 57 FR 7163, February 28, 1992.
- 11. National Institute for Occupational Safety and Health, (NIOSH), Rockville, MD. Explosive azide hazard, Publication No. 78-127, 1976.

2°C 8°C	Temperature Limitation	IVD	In Vitro Diagnostic Medical Device
$\boxtimes$	Use By Date	LOT	Batch Code
NON	Non-Sterile	[]i	Consult Instructions for Use
EC REP	Representative in the European Community		Manufacturer

© 2020, BioGenex Laboratories. All rights reserved.

Category	Antibodies	Revision No.	J
Document No.	932-073M-EN	Release Date	11-May-2022