

# Anti-P504S (AMACR) [13H4]

Catalog No.	Description	
AN449-5M	6 ml of Ready-to-Use Antibody for use with BioGenex Super Sensitive TM Detection Systems OR equivalent detection system	
AN449-10M	10 ml of Ready-to-Use Antibody in a barcode labeled vial for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems and i6000 <sup>TM</sup> Automated Staining Systems	
NU449-UC	1 ml of Concentrated Antibody for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems OR equivalent detection system	
NU449-5UC	0.5 ml of Concentrated Antibody for use with BioGenex Super Sensitive <sup>TM</sup> Detection Systems OR equivalent detection system	
AY449-YCD	Ready-to-Use Antibody in Barcode labeled vial for use on the Xmatrx <sup>®</sup> Elite Staining System, 160 tests	
AY449-50D	Ready-to-Use Antibody in Barcode labeled vial for use on the Xmatrx <sup>®</sup> Elite Staining System, 50 tests	

Clone	Species	Ig Class
13H4	Rabbit	IgG

# **Intended Use**

**For In Vitro Diagnostic Use.** This antibody is designed for the specific localization of P504S in formalin-fixed, paraffinembedded (FFPE) tissue sections. Evaluation must be performed by a qualified pathologist.

# **Summary and Explanation**

P504S is a gene that encodes a protein Alpha-Methylacyl-Coenzyme A Racemase that is involved in the metabolism of branched-chain fatty acid and bile acid intermediates. P504S antibody stains human Alpha Methylacyl CoA Racemase in the cytoplasm of target prostatic cells.

## **Storage and Handling**

**Store at 2-8°C.** Fresh dilutions, if required, should be prepared prior to use and are stable and steady for up to one day at room temperature (20-26°C). Diluted antibody preparations can be refrigerated or frozen for extended shelf life.

# **Principles of the Procedure**

Antigen detection by immunohistochemistry (<u>IHC</u>) is a two-step process wherein the primary antibody binds to the antigen of interest and that binding is detected by a chromogen. The <u>primary</u> antibody may be used in IHC using manual techniques or BioGenex

ECIREPI

Emergo Europe, Prinsessegracht 20, 2514AP The Hague, The Netherlands

Automated Staining System. Positive and negative controls should always be run simultaneously with all patient specimens.

## **Reagents Provided**

Rabbit Monoclonal Antibody P504S is affinity purified and diluted in PBS, pH 7.2, containing 1% BSA and 0.09% sodium azide.

## **Dilution of Primary Antibody**

BioGenex Ready-to-Use antibodies have been optimized for use with the recommended BioGenex Detection System and should not require further dilution.

BioGenex concentrated antibodies must be diluted in accordance with the recommended protocol when used with the recommended BioGenex Detection System.

# **Recommended Protocol**

Refer to the following table for conditions specifically recommended for this antibody. Refer to the BioGenex website for guidance on specific staining protocols or other requirements.

	P: C	
Parameter	BioGenex	
1 at affected	Recommendations	
Control Tissue	Prostate Carcinoma tissue as available with Biogenex FB- 449N* & FG-449N*	
Recommended Dilution for Concentrated Antibody	1:20-50 in HK941	
Recommended Pretreatment (Manual/i6000)**	EZ-AR2 (HK522-XAK)	
Recommended Pretreatment (Xmatrx)	EZ-AR2 Elegance (HX032-YCD & HX046- 08XN)	
Antibody Incubation (Manual/i6000)	30-60 Min at RT	
Antibody Incubation (Xmatrx)	30-60 Min at 25°C	
	Use BioGenex Two-Step <b>OR</b>	
Detection System for	One-Step Super Sensitive <sup>TM</sup>	
Manual, Xmatrx & i6000	Polymer-HRP IHC Detection	
systems***	System/DAB; see p. 2 for more information	

<sup>\*</sup>FB: positive control micro chamber slides, FG: positive control microscopic slides. Xmatrx requires micro chamber slides.

\*\*Pretreatment times will vary based on individual microwave power.

\*\*\*For automation systems (Xmatrx-Elite & i6000 Diagnostics), refer to the factory protocols provided with the instrument.

CategoryAntibodiesRevision No.KDocument No.932-449NRelease Date22-Jul-2022



Emergo Europe, Prinsess	segracht 20, 2514AP The Hague, The Netherlands
EC REP	

Detection	Two-Step	One-Step	Link and
System	HRP Kit	HRP Kit	Label Kit
Manual	QD440-XAKEN (1000 Test) QD430-XAKEN (1000 Test)	QD630-XAKEN (1000 Test)	QP300- XAKE (1000 Test)
Manual	QD420-YIKEN (500 Test) QD400-60KEN (60 Test)	QD620-XAKEN (500 Test)	QP900- 9LE (500 Test)
Xmatrx - Automation	QD550-YCDEN (200 Test)	QD610-YADEN (200 Test)	N/A
i6000 - Automation	QD410-YAXEN (200 Test)	QD610-YAXEN (200 Test)	N/A
For more information, visit www.biogenex.com.			

#### **Precautions**

This product contains sodium azide at concentrations of less than 0.1%. Sodium azide is not classified as a hazardous chemical at the product concentrations, but proper handling protocols should be observed. For more information, a Safety Data Sheet (SDS) for sodium azide is available upon request. Dispose of unused reagents according to Local, State and Federal Regulations. Wear suitable Personal Protective Equipment, do not pipette reagents by mouth, and avoid contact of reagents and specimens with skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with copious amounts of water.

# **Quality Control**

Refer to BioGenex detection system documents for guidance on general quality control procedures.

#### **Troubleshooting**

Refer to the troubleshooting section in the documentation for BioGenex Detection Systems (or equivalent detection systems) for remedial actions on detection system related issues, or contact BioGenex Technical Support Department at 1-800-421-4149 or <a href="mailto:support@biogenex.com">support@biogenex.com</a> or your local distributor to report unusual staining.

### **Expected Results**

This antibody stains cytoplasm in positive cells in formalin-fixed, paraffin embedded tissue sections. An example image of a tissue section stained with this antibody can be found on the product page on the BioGenex website. Interpretation of the staining result is solely the responsibility of the user. Experimental results should be confirmed by a medically-established diagnostic product or procedure.

### **Limitations of the Procedure**

Improper tissue handling and processing prior to immunostaining can lead to inconsistent results. Variations in embedding and fixation or the nature of the tissue may lead to variations in results.

Category	Antibodies	Revision No.	K
Document No.	932-449N	Release Date	22-Jul-2022

Endogenous peroxidase activity or pseudo peroxidase activity in erythrocytes and tissue biotin may result in non-specific staining based on the detection system employed. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive with horseradish peroxidase systems. Improper counterstaining and mounting may compromise the interpretation of results.

# **Bibliography**

- 1. Xu J, Stolk JA, Zhang X et al. Cancer Res 2000; 60:1677-1682.
- 2. Jiang Z, Woda BA, Rock KL et al. Am J Surg Pathol 2001; 25:1397-1404

2°C 8°C	Temperature Limitation	IVD	In Vitro Diagnostic Medical Device
$\boxtimes$	Use By Date	LOT	Batch Code
NON	Non-Sterile	[]i	Consult Instructions for Use
EC REP	Representative in the European Community		Manufacturer

© 2020, BioGenex Laboratories. All rights reserved.