

Emergo Europe, Prinsessegracht 20, 2514AP The Hague, The Netherlands

Anti-TTF-1 [NX2.1/690]

Catalog No.	Description		
AMA25-5M	6 ml of Ready-to-Use Antibody for use with BioGenex Super Sensitive TM Detection Systems OR equivalent detection system		
AMA25-10M	10 ml of Ready-to-Use Antibody in abarcode labeled vial for use with BioGenexSuper Sensitive TM Detection Systems andi6000 TM Automated Staining Systems		
MUA25-UC	1 ml of Concentrated Antibody for use with BioGenex Super Sensitive TM Detection Systems OR equivalent detection system		
MUA25-5UC	0.5 ml of Concentrated Antibody for use		
AXA25-YCD	Ready-to-Use Antibody in Barcode labeled		
AXA25-50D	Ready-to-Use Antibody in Barcode labeled vial for use on the Xmatrx [®] Elite/Ultra Staining System, 50 tests		

Clone	Species	Ig Class
NX2.1/690	Mouse	IgG2b, kappa

Intended Use

For In Vitro Diagnostic Use. This antibody is designed for the specific localization of TTF-1 in formalin-fixed, paraffinembedded (FFPE) tissue sections. Evaluation must be performed by a qualified pathologist.

Summary and Explanation

Thyroid Transcription Factor-1 (TTF-1), also known as thyroidspecific enhancer-binding protein (T/EBP), is a 40 kD protein that is a member of NKx2 family of homeodomain transcription factors that regulates the expression of thyroid- and lung-specific genes. It is a very selective marker for adenocarcinomas of lung and thyroid origin. Nuclear localization of this protein is seen in the epithelial cells of thyroid gland and lung. The anti-TTF-1 antibody is a useful tool for differentiating pulmonary adenocarcinoma from metastatic breast carcinoma and mesothelioma.

Storage and Handling

Store at 2-8°C. Fresh dilutions, if required, should be prepared prior to use and are stable and steady for up to one day at room temperature (20-26°C). Diluted antibody preparations can be refrigerated or frozen for extended shelf life.

Principles of the Procedure

Antigen detection by immunohistochemistry (IHC) is a two-step process wherein the primary antibody binds to the antigen of interest and that binding is detected by a chromogen. The primary antibody may be used in IHC using manual techniques or BioGenex Automated Staining System. Positive and negative controls should always be run simultaneously with all patient specimens.

Reagents Provided

Mouse Monoclonal Antibody TTF-1 is affinity purified and diluted in PBS, pH 7.2, containing 1% BSA and 0.09% sodium azide.

Dilution of Primary Antibody

BioGenex Ready-to-Use antibodies have been optimized for use with the recommended BioGenex Detection System and should not require further dilution.

BioGenex concentrated antibodies must be diluted in accordance with the recommended protocol when used with the recommended BioGenex Detection System.

Recommended Protocol

Refer to the following table for conditions specifically recommended for this antibody. Refer to the BioGenex website for guidance on specific staining protocols or other requirements.

Parameter	BioGenex Recommendations
Control Tissue	Lung cancer as available with BiogenexFB-A25M* & FG-A25M*
Recommended Dilution for Concentrated Antibody	1:10-20 in HK941
Recommended Pretreatment (Manual/i6000)**	EZ-AR2 (HK522-XAK)
Recommended Pretreatment (Xmatrx)	EZ-AR2 Elegance (HX032-YCD)
Antibody Incubation (Manual/i6000)	30-60 min at RT
Antibody Incubation (Xmatrx)	30-60 min at 25°C
Detection System for Manual, Xmatrx & i6000 systems***	Use BioGenex Two-Step OR One-Step Super Sensitive TM Polymer-HRP IHC Detection System/DAB; see p. 2 for more information

*FB: positive control barrier slides, FG: positive control nonbarrier slides. Xmatrx requires barrier slides.

Pretreatment times will vary based on individual microwave power. *For automation systems (Xmatrx-Elite, Xmatrx-Ultra & i6000 Diagnostics), refer to the factory protocols provided with the instrument.

Category	Antibodies	Revision No.	Е
Document No.	932-A25M-EN	Release Date	June 01, 2021



CE	Emergo Europe, Prinsessegracht 20, 2514AP The Hague, The Netherlands
EC RE	P

Detection	Two-Step	One-Step	Link and
System	HRP Kit	HRP Kit	Label Kit
Manual	QD440-XAKE (1000 Test) QD430-XAKE (1000 Test) QD420-YIKE (500 Test) QD400-60KE	QD630-XAKE (1000 Test) QD620-XAKE (500 Test)	QP300-XAKE (1000 Test) QP900-9LE (500 Test)
	(60 Test)		
Xmatrx - Automation	QD550-YCDE (200 Test)	QD610-YADE (200 Test)	N/A
i6000 - Automation	QD410-YAXE (200 Test)	QD610-YAXE (200 Test)	N/A
For more information, visit <u>www.biogenex.com</u> .			

Precautions

This product contains sodium azide at concentrations of less than 0.1%. Sodium azide is not classified as a hazardous chemical at the product concentrations, but proper handling protocols should be observed. For more information, a Safety Data Sheet (SDS) for sodium azide is available upon request. Dispose of unused reagents according to Local, State and Federal Regulations. Wear suitable Personal Protective Equipment, do not pipette reagents by mouth, and avoid contact of reagents and specimens with skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with copious amounts of water.

Quality Control

Refer to BioGenex detection system documents for guidance on general quality control procedures.

Troubleshooting

Refer to the troubleshooting section in the documentation for BioGenex Detection Systems (or equivalent detection systems) for remedial actions on detection system related issues, or contact BioGenex Technical Support Department at 1-800-421-4149 or <u>support@biogenex.com</u> or your local distributor to report unusual staining.

Expected Results


This antibody stains nucleus in positive cells in formalin-fixed, paraffin embedded tissue sections. An example image of a tissue section stained with this antibody can be found on the product page on the BioGenex website. Interpretation of the staining result is solely the responsibility of the user. Experimental results should be confirmed by a medically-established diagnostic product or procedure.

Limitations of the Procedure

Improper tissue handling and processing prior to immunostaining can lead to inconsistent results. Variations in embedding and fixation or the nature of the tissue may lead to variations in results. Endogenous peroxidase activity or pseudo peroxidase activity in erythrocytes and tissue biotin may result in non-specific staining based on the detection system employed. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive with horseradish peroxidase systems. Improper counterstaining and mounting may compromise the interpretation of results.

Bibliography

- 1. Nakamura K, et al. Brain Res Dev Res 130:159-66, 2001.
- 2. Hecht JL, et al. Am J ClinPathol. 116:483-8, 2001.
- 3. Holzinger, et al. Hybridoma, 15:49-53, 1996.
- 4. Bejarano PA, et al. Mod Pathol 9:445-52,1996.
- 5. Di Loreto C, et al. Cancer Lett 124:73-8, 1998.
- Center for Disease Control. Decontamination of Laboratory Sink Drains to Remove Azide Salts.Center for Disease Control Manual Guide—Safety Management, No. CDC-22, Atlanta, Georgia. April 30, 1976.
- 7. Kiernan JA. Histological and Histochemical Methods: Theory and Practice. New York: Pergamon Press 1981.
- 8. Nadji M, Morales AR. Immunoperoxidase, part 1: the techniques and its pitfall. Lab Med 1983; 14:767-770.
- Omata M, Liew CT, Ashcavai M, Peters RI. Nonimmunologic binding of horseradish peroxidase to hepatitis B surface antigen. A possible source of error in immunohistochemistry.Am J ClinPathol. May, 1980;73(5):626-632.
- U.S. Congress. Clinical Laboratory Improvement Amendments of 1988: Final Rule, 57 FR 7163, February 28, 1992.
- National Institute for Occupational Safety and Health, (NIOSH), Rockville, MD. Explosive azide hazard, Publication No. 78-127, 1976.



© 2020, BioGenex Laboratories. All rights reserved.

Category	Antibodies	Revision No.	Е
Document No.	932-A25M-EN	Release Date	June 01, 2021