# **BioGenex**

**48810 Kato Road, Suite 100E & 200E** Fremont, CA 94538 Tel : +1 (800) 421-4149, Fax: +1 (510) 824-1490, support@biogenex.com



# DATA SHEET eFISH COL1A1 Dual Color Break Apart Probe

# Catalog No. FP054-10XE- 100µl-10 test FP054-20XE- 200µl-20 test

Doc No: 932-FP054E Rev: D Date of Release: 05-Aug-2020

Material Provided: One vial of eFISH probe in hybridization buffer (RTU).

## **Recommended detection system (Not supplied):**

Either of the following detection system is recommended depending on the automation/manual platform used:

eFISH Kit	Cat #	Description
eFISH Histo	DF-500-20XE	Automation
eFISH Cyto	DF-510-20XE	Automation

#### **Intended Use:**

The BioGenex eFISH COL1A1 Dual Color Break Apart Probe is designed to detect translocations involving COL1A1gene at 17q21 in formalin-fixed, paraffin-embedded tissue or cells by fluorescence in situ hybridization (FISH).

BioGenex eFISH COL1A1 Dual Color Break Apart Probe comes in hybridization buffer. The probe contains green-labeled polynucleotides (Green: excitation at 503 nm and emission at 528 nm, similar to FITC) targeting sequence in 17q21 proximal to the COL1A1 gene andOrange labeled polynucleotides (Orange: excitation at 547 nm and emission at 572 nm, similar to rhodamine), which target sequences mapping mapping in 17q21 distal to the COL1A1 gene.

## **Summary and Explanation**

Fluorescence *in situ* hybridization (FISH) is a robust technique of cytogenetic used for the detection of chromosomal aberrations, presence or absence of specific DNA sequence in native context. In this technique florescent probes bind to the target sequence of DNA in chromosome. High specificity and sensitivity coupled rapid and an accurate result has proven role of FISH in both research and diagnosis of solid tumor and hematological malignancies. As technique of cancer cytogenetics, FISH, can be used to identify genetic aberrations viz., deletions, amplification and translocation in tissue sections or within individual cells. FISH is also used for use in genetic counseling, medicine, and species identification. FISH can also be used to detect and localize specific RNA targets in cells,

circulating tumor cells, and tissue samples<sup>1,2,3,4,5</sup>.

In FISH procedure, fixed tissue sections are pretreated to expose target DNA or mRNA sequences. An appropriately labeled probe is hybridized to the exposed target DNA or mRNA sequences in the cells. Subsequent stringent washing steps remove any probe that is non-specifically bound to the tissue section. Subsequently slides are mounted using DAPI/antifade and can be visualized under fluorescence microscope using appropriate filter set.

# **Principles of the Procedure**

*In Situ* hybridization (ISH) allows the detection and localization of definitive nucleic acid sequences directly within a cell or tissue. High specificity is ensured through the action of annealing of fluorescence probe nucleic acid sequence to complementary target nucleic acid sequence. ISH techniques can be used to identify genetic aberrations like deletions, amplification, and translocationin tissue sections or within individual cells.

# **Storage and Handling**

BioGenex eFISH COL1A1 Dual Color Break Apart Probe must be stored at 2-8°C protected from light and is stable through the expiry date printed on the label.

# **Specimen Collection and Slide Preparation**

Tissues fixed in 10% (v/v) formalin are suitable for use prior to paraffin embedding and sectioning.

## **FISH Staining procedure**

- (a) The BioGenex eFISH probes are supplied in hybridization buffer and used without further dilution.
- (b) Protocol:

Please refer to the eFISH probe specific instruction/protocol for automated or semiautomated FISH processing platform (Xmatrx<sup>®</sup>-Infinity, Xmatrx<sup>®</sup>-Nano and Xmatrx<sup>®</sup>mini.

Further processing, such as washing and counter-staining, can becompleted according to the user's needs. For a particularly user-friendlyperformance, we recommend the use of a BioGenexeFISH kit.

These systems were also used for the confirmation of appropriateness of the eFISH COL1A1 Dual Color Break Apart Probe.

**Disclaimer**: The above information is provided for reference only. Each end-user is responsible for developingand validating optimal testing conditions for use with this product.

# **BioGenex\_**

₩ 48810 Kato Road, Suite 100E & 200E Fremont, CA 94538 Tel : +1 (800) 421-4149. Fax: +1 (510) 824-1490, support@biogenex.com



Emergo Europe, Prinsessegracht 20 2514 AP The Hague, The Netherlands

# Troubleshooting

Contact BioGenex Technical Service Department at **1-800-421-4149 or your local distributor** to report unusual staining.

#### **Expected Results**

The use of eFISH PDGFB Dual Color Break Apart Probe along with appropriate filters produces orange and green signal for hybridization regions of 17q21. Normal interphase cells or cells without 17q21translocation, two orange/two green fusion signals appear. In regions where 17q21 is affected by translocation, one separate green signal and one separate orange signal are observed.

However, we recommend the use of a control sample in which the translocation status of 17q21is known to judge the specificity of the signals with each hybridization reaction.

Care should be taken not to evaluate overlapping cells, in order to avoid false results, e.g. an amplification of genes. Due to decondensed chromatin, single FISH signals can appear as small signal clusters. Thus, two or three signals of the same size, separated by a distance equal to or less than the diameter of one signal, should be counted as one signal.

## Limitations of the Procedure

Correct treatment of tissues prior to and during fixation, embedding, and sectioning is important for obtaining optimal results. Inconsistent results may be due to variations in tissue processing, as well as inherent variations in tissue. The results from *in situ* hybridization must be correlated with other laboratory findings.

## Bibliography

- 1. Gall, J. G. and Pardue, M. L. (1969). Proc. Natl. Acad. Sci. USA63, 378 -383.
- 2. Rudkin, G. T. and Stollar, B. D. (1977). Nature 265,472-473.
- 3. Hougaard, D. M., Hansen, H. and Larsson, L. I. (1997). Histochem. Cell Biol. 108,335 -344.
- 4. Bauman, J. G., Wiegant, J., Borst, P. and van Duijn, P. (1980). Cell Res. 128,485 490.
- **5.** O'Connor et al. (2008). *Nature Education* 1(1):171.
- 6. Kievits T, et al. (1990) Cytogenet Cell Genet 53: 134-6.
- 7. Labropoulos SV, Razis ED (2007) Biologics 4: 347-53.
- 8. Patel KU, et al. (2008) Human Pathol 39: 184-93.
- 9. Shimizu A, et al. (1999) Cancer Res 59: 3719-23.

2°C	Temperature Limitation	IVD	In Vitro Diagnostic Medical Device
	Use By Date	LOT	Batch Code
NON STERLE	Non-Sterile		Consult Instructions for Use
EC REP	Representative in the European Community		BioGenex