BioGenex 48810 Kato Road, Suite 100E & 200E Fremont, CA 94538 Tel : +1 (800) 421-4149, Fax: +1 (510) 824-1490, support@biogenex.com

Data Sheet eFISH Probes for Chromosome Enumeration

Background


Chromosome Enumeration Probes are designed for identification and enumeration of human chromosomes in interphase cells and as an adjunct to standard karyotyping in metaphases. These probes will produce sharp, bright signals specific for each individual chromosome.

References

Probe Description


For most chromosomes, direct labeled BioGenex CEN Probes hybridizing to highly repetitive human satellite DNA sequences mainly located at the centromeric regions of chromosomes are applicable. As several chromosomes share the same repetitive sequences resulting in cross-hybridization signals, they cannot be differentiated by centromere specific probes. Instead, these chromosomes can be identified by direct labeled BioGenex eFISH Probes hybridizing in close proximity to the respective satellite DNA sequences or to other chromosome specific loci.

The BioGenex eFISH 1p12 Probe is designed to hybridize in close proximity of centromere 1 at 1p12 harboring WARS2, and HAO2. Since chromosomes 1, 5, and 19 share the same repetitive sequences, they cannot be differentiated by probes detecting centromere specific repeats.




SPEC 1p12 probe map (not to scale)

The BioGenex eFISH 2q11 Probe is specific for the AFF3 (AF4/FMR2 family, member 3) gene region in 2q11.2. Due to cross-hybridizations of chromosome 2 alpha satellites to other centromeric regions, probes specific for 2q11 are frequently used for chromosome 2 copy number detection.



SPEC 2q11 probe map (not to scale)


The BioGenex eFISH 4p11 Probe is designed to hybridize in close proximity of centromere 4 at 4p11 harboring the ZAR1 (zygote arrest 1)gene. For an unambiguous enumeration of chromosome 4 the SPEC 4p11 is found to be more suitable.



SPEC 4p11 probe map (not to scale)

The BioGenex eFISH13q12 Probe is

designed to hybridize in close proximity of centromere 13 at 13q12. Since chromosomes 13 and 21 share the same repetitive sequences, they cannot be differentiated by probes detecting centromere specific repeats.



SPEC 13q12 probe map (not to scale)

The ZytoLight [®] SPEC 21q22 Probe hybridizesto the so-called Down Syndrome CriticalRegion on 21q22 commonly duplicated incases with partial trisomy 21.

Since chromosomes 13 and 21 share thesame repetitive sequences, they cannotbe differentiated by probesdetecting centromere specific repeats.



SPEC 21q22 probe map (not to scale)

Result

In a normal interphase nucleus, two signals are expected using Chromosome Enumeration Probes specific for autosomes. Using chromosome Y specific probes will result in normal male cells in one signal and in normal female cells in no signal. Using chromosome X specific probes will

Doc No: 932-FP084-102; Rev: D Effc Date: 10-Aug-2020

Bacommonded detection suctom (Net cumplied)

BioGenex 48810 Kato Road, Suite 100E & 200E Fremont, CA 94538 Tel : +1 (800) 421-4149, Fax: +1 (510) 824-1490, support@biogenex.com

result in normal male cells in one signal and in normal female cells in two signals per nucleus. Other signal patterns indicate numerical aberrations of the

respective chromosome

| Sr# | Enumeration Probes | Cat # | Color |
|-----|----------------------|---------------|--------|
| 1 | eFISH 1p12 Probe | FP084-10X/20X | Orange |
| 2 | eFISH 2q11 Probe | FP085-10X/20X | Orange |
| 3 | eFISHCEN 3 Probe | FP086-10X/20X | Orange |
| 4 | eFISH 4p11 Probe | FP087-10X/20X | Orange |
| 5 | eFISHCEN 6 Probe | FP088-10X/20X | Green |
| 6 | eFISH CEN 7 Probe | FP089-10X/20X | Orange |
| 7 | eFISHCEN 8 Probe | FP090-10X/20X | Green |
| 8 | eFISH CEN 9 Probe | FP091-10X/20X | Orange |
| 9 | eFISHCEN 10 Probe | FP092-10X/20X | Orange |
| 10 | eFISH CEN 11 Probe | FP093-10X/20X | Green |
| 11 | eFISHCEN 12 Probe | FP094-10X/20X | Orange |
| 12 | eFISH 13q12 Probe | FP095-10X/20X | Orange |
| 14 | eFISH CEN 17 Probe | FP097-10X/20X | Orange |
| 15 | eFISHCEN 18 Probe | FP098-10X/20X | Orange |
| 16 | eFISH 21q22 Probe | FP099-10X/20X | Orange |
| 17 | eFISH CEN X Probe | FP100-10X/20X | Orange |
| 18 | eFISH CEN Yq12 Probe | FP101-10X/20X | Green |

Doc No: 932-FP084-102; Rev: D Effc Date: 10-Aug-2020

٦

Decommanded detection system (Net symplical)

Bibliography

- 1. Gall, J. G. and Pardue, M. L. (1969). Proc. Natl. Acad. Sci. USA63, 378 -383.
- 2. Rudkin, G. T. and Stollar, B. D. (1977). Nature 265,472-473.
- 3. Hougaard, D. M., Hansen, H. and Larsson, L. I. (1997). Histochem. Cell Biol. 108,335 -344.
- 4. Bauman, J. G., Wiegant, J., Borst, P. and van Duijn, P. (1980). *Cell Res.* 128,485 -490.
- 5. O'Connor et al. (2008). Nature Education 1(1):171.
- 6. Kievits, T. et al. Cytogenet. Cell Genet. 53, 134–136 (1990).
- 7. Antonescu C. (2000).J MolDiagn 2000;2:132-8
- 8. Germano G, et al.(2010) Cancer Res 70: 2235-44.
- 9. Kuroda M, et al. (1995). Am J Pathol. Nov;147(5):1221–1227

Doc No: 932-FP084-102; Rev: D Effc Date: 10-Aug-2020