

DATA SHEET
Hsa-miR-222 fluorescened oligo probe

Catalog No. HM222-100	Description One vial of 0.650 ml of probe in hybridization buffer
--	---

Analyte Specific Reagent. Analytical and performance characteristics are not established.

Doc. No. 932-HM222-100 Rev:C

Date of release: 14-Aug-2020

Description

The Hsa-miR-222 probe has been designed from mature human miR-222 sequence. This fluorescened probe is provided in a hybridization buffer for localization of miRNA in FFPE tissue by *In Situ* hybridization.

Specifications

The Hsa- miR-222 identifies mature miR-222 sequences in formalin-fixed, paraffin-embedded human tissues and freshly prepared frozen tissues by *in situ* hybridization. This probe does not react with normal human mRNA or nuclear DNA present in tissues.

Storage and Handling

Store the reagent at 2-8 °C. Do not freeze. Do not use after expiration date on vial. The reagent must be brought to room temperature before use. (Important! The presence of precipitates induces background staining).

Precautions:

For professional use the probe contains formamide. Formamide is classified as a teratogen. Pregnant workers should keep exposure to a minimum. Avoid inhalation, ingestion, and contact with unprotected skin. If skin contact occurs, wash thoroughly with soap and water. For more information, refer to the Material Safety Data Sheet, which is available upon request.

Quality Control

Each lot of this micro RNA probe is tested by *In Situ* hybridization for Quality Control purposes. Refer to the BioGenex Quality Control Testing Conditions table for additional information.

References

1. Qian K et al. Hsa-miR-222 is involved in differentiation of endometrial stromal cells in vitro. *Endocrinology*. 150(10):4734-43(2009).
2. Kloosterman WP et al. *in situ* detection of miRNAs in animal embryos using LNA-modified oligonucleotide probes. *Nature Methods*, 3, 27 – 29 (2006).
3. Wheeler G. et al. *In situ* detection of animal and plant microRNAs. *DNA Cell Biol*, 26, 251–255 (2007).
4. Nuovo GJ. In situ detection of precursor and mature microRNAs in paraffin embedded, formalin fixed tissues and cell preparations. *Methods* 44(1), 39–46 (2008).
5. Song R. et al. *In situ* hybridization detection of microRNAs. *Methods Mol Biol*. 629, 287-94 (2010)

BioGenex Quality Control Testing Conditions

Parameter	Conditions used
Control Tissue	.BREAST CARCINOMA/ LUNG CARCINOMA (FB-HM222).
Tissue Type	Formalin-fixed, paraffin-embedded cancer tissues