

DATA SHEET
Hsa-miR-337 fluoresceinated oligo probe**Catalog No.**
HM337-100**Description**
One vial of 0.650 ml of probe in hybridization buffer**Analyte Specific Reagent. Analytical and performance characteristics are not established.**

Doc. No. 932-HM337-100

Rev. B

Date of release: 17-Aug-2020

Description

The Hsa-miR-337 probe has been designed from mature human miR-337 sequence. This fluoresceinated probe is provided in a hybridization buffer for localization of miRNA in FFPE tissue by *In-Situ* hybridization.

Specifications

The Hsa-miR-337 identifies mature miR-337 sequences in formalin-fixed, paraffin-embedded human tissues and/or freshly prepared frozen tissues by *in situ* hybridization. This probe does not react with normal human mRNA or nuclear DNA present in tissues.

Storage and Handling

Store the reagent at 2-8 °C. Do not freeze. Do not use the reagent after expiration date on vial. The reagent must be brought to room temperature before use. (Important! The presence of precipitates induces background staining).

Precautions:

For professional use the probe contains formamide. Formamide is classified as a teratogen. Pregnant workers should keep exposure to a minimum. Avoid inhalation, ingestion, and contact with unprotected skin. If skin contact occurs, wash thoroughly with soap and water. For more information, refer to the Material Safety Data Sheet, which is available upon request.

Quality Control

Each lot of this miRNA probe is tested by *In Situ* hybridization for Quality Control purposes. Refer to the BioGenex Quality Control Testing Conditions table for additional information.

References

1. Plummer PN, et al.. (2013). MicroRNAs regulate tumor angiogenesis modulated by endothelial progenitor cells. *Cancer Res.* 73(1), 341–52.
2. Palmieri A, et al.. (2008). PerioGlas regulates osteoblast RNA interfering. *J Prosthodont.* 17(7), 522–6.
3. Annalisa P, et al.. (2008). Anorganic bovine bone and a silicate-based synthetic bone activate different microRNAs. *J Oral Sci.* 50(3), 301–7.
4. Zhang R, et al.. (2014). Levels of HOXB7 and miR-337 in pancreatic ductal adenocarcinoma patients. *Diagn Pathol.* 18(9): 61.

BioGenex Quality Control Testing Conditions

Parameter	Conditions used
Control Tissue	Human Lymph Node tissues
Tissue Type	Formalin-fixed, paraffin-embedded normal tissues